

04.01-06/07/95-01408



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

REGION IV

345 COURTLAND STREET, N.E.
ATLANTA, GEORGIA 30365

June 7, 1995

4WD-FFB

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

Ms. Katherine Landman
Department of the Navy - Atlantic Division
Naval Facilities Engineering Command
Code 1823
Norfolk, Virginia 23511-6287

SUBJ: MCB Camp Lejeune
Draft Treatability Work Plan
Operable Unit No. 14 - Site 69

Dear Ms. Landman:

The enclosed comments were just recently received from the "Laboratory Evaluation & Quality Assurance Section" located in the Athens laboratory. The comments appear to be minor and can be corrected in the next submission.

If you have any questions or comments, please call me at (404) 347-3016 or voice mail, (404) 347-3555, x-6459.

Sincerely,

A handwritten signature in cursive script that reads "Gena D. Townsend".

Gena D. Townsend
Senior Project Manager

Enclosure

cc: Patrick Waters, NCDEHNR
Neal Paul, MCB Camp Lejeune

Comments

1. Section 6, Table 6-1
 - a. In order to be complete, the table should include the extraction/preparation methods used in analysis of the samples.
 - b. The table should reference the source of the analytical methods. In addition, periodic revisions are made to SW-846. If a particular method has been revised, the updated version should be used. The reference should indicate from which edition or revision the methods are taken.
2. Section 6.1.2.3 - Region IV policy is not to filter water samples for inorganics (metals) analyses. If results for filtered samples are desired, these must be in addition to unfiltered.
3. Sections 7.2.1, 7.3.2 and 7.3.3 - Quality control samples should be subject to the same chemical preservation as used for the associated samples. That is, trip blanks should be preserved with 1:1 HCl, and equipment and field blanks should be preserved with HNO₃.